

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 002358  
11:43 30/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:25  
PRESSURIZING 1:29  
STERILIZING 2:28  
136.4 C 220kPa 2:28  
136.2 C 218kPa 2:58  
136.2 C 222kPa 3:28  
136.4 C 222kPa 3:58  
136.4 C 222kPa 4:28  
136.2 C 222kPa 4:58  
136.2 C 222kPa 5:28  
136.4 C 219kPa 5:58  
136.4 C 219kPa 5:58  
VENTING 5:58  
AIR DRYING 6:48  
CYCLE COMPLETE 8:18

-----  
Digital Signature #  
D35C12F6C9C3AC19  
-----