

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2267

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2022 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.5 F 0.0  
2:00 170.6 F 1.7  
4:00 194.7 F 6.4  
6:00 214.1 F 11.7  
8:00 224.8 F 7.6  
10:00 239.1 F 13.4  
12:00 239.7 F 11.3  
14:00 238.9 F 10.2  
16:00 253.5 F 17.4  
18:00 265.9 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.1 F 28.9  
2:00 272.5 F 29.4  
3:00 271.4 F 28.6  
4:00 272.4 F 29.2  
  
Min 271.4 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 223.6 F 0.0  
10:00 226.7 F 0.0  
15:00 226.9 F 0.0  
20:00 229.5 F 0.0  
25:00 233.5 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

08/30/2022 02:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_