
Midmark M11 - v1.0.5
Total Cycles: 3291

Name: RITTER2
Sterilizer ID: V2313228
01/11/2024 03:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.1 F 0.0
2:00 132.2 F 1.0
4:00 157.9 F 3.7
6:00 183.4 F 7.6
8:00 216.6 F 11.8
10:00 217.7 F 5.1
12:00 230.1 F 9.2
14:00 237.8 F 11.3
16:00 238.9 F 11.1
18:00 237.5 F 9.7
20:00 251.1 F 16.3
22:00 262.6 F 22.9
24:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.2
2:00 272.1 F 29.5
3:00 272.3 F 29.1
4:00 271.9 F 29.5

Min 271.4 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 236.8 F 0.0
10:00 226.6 F 0.0
15:00 230.9 F 0.0
20:00 227.8 F 0.0
25:00 208.6 F 0.0
30:00 192.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:15
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

01/11/2024 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____