
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001965
10:46 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:52
STERILIZING 2:56
136.5 C 225kPa 2:56
136.4 C 227kPa 3:26
136.4 C 224kPa 3:56
136.4 C 225kPa 4:26
136.3 C 226kPa 4:56
136.4 C 223kPa 5:26
136.2 C 224kPa 5:56
136.0 C 223kPa 6:26
136.0 C 223kPa 6:26
VENTING 6:26
AIR DRYING 7:16
CYCLE COMPLETE 37:16

Digital Signature #
988B2393862F6280
