
Midmark M11 - v1.0.0
Total Cycles: 1458

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 189.9 F 2.7
4:00 185.3 F 6.5
6:00 211.6 F 12.1
8:00 224.9 F 6.8
10:00 234.5 F 12.2
12:00 235.2 F 9.0
14:00 240.7 F 9.8
16:00 246.5 F 13.5
18:00 258.9 F 20.1
20:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.2
2:00 272.4 F 28.9
3:00 271.8 F 28.4
4:00 271.3 F 28.5

Min 271.2 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 227.6 F 0.0
10:00 233.1 F 0.0
15:00 228.5 F 0.0
20:00 229.5 F 0.0
25:00 227.7 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

12/15/2021 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____