

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4103

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2023 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
2:00 173.8 F 3.0  
4:00 186.2 F 7.0  
6:00 206.4 F 11.6  
8:00 217.0 F 5.0  
10:00 228.1 F 9.7  
12:00 236.7 F 9.4  
14:00 241.9 F 11.7  
16:00 239.4 F 10.1  
18:00 252.3 F 16.4  
20:00 263.3 F 22.7  
22:00 272.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.9 F 29.2  
2:00 271.6 F 28.5  
3:00 272.3 F 28.9  
4:00 272.0 F 29.0

Min 271.0 F 28.1  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
5:00 182.8 F 0.0  
10:00 197.5 F 0.0  
15:00 196.8 F 0.0  
20:00 195.5 F 0.0  
25:00 192.4 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 22:18  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

12/11/2023 12:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_