
Midmark M11 - v1.0.0
Total Cycles: 1916

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.4 F 0.0
2:00 154.9 F 1.7
4:00 176.8 F 5.2
6:00 201.5 F 9.9
8:00 215.0 F 5.5
10:00 228.5 F 9.0
12:00 237.4 F 10.9
14:00 240.6 F 11.3
16:00 238.0 F 9.5
18:00 252.2 F 16.4
20:00 264.0 F 23.3
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 28.9
2:00 271.7 F 28.5
3:00 272.7 F 28.9
4:00 271.6 F 28.2

Min 271.6 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 177.6 F 0.0
10:00 179.3 F 0.0
15:00 172.6 F 0.0
20:00 171.9 F 0.0
25:00 169.4 F 0.0
30:00 167.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:00
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

05/05/2022 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____