

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000324  
9:48 11/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:17  
STERILIZING 4:28  
136.0 C 227kPa 4:28  
136.4 C 233kPa 4:58  
136.4 C 233kPa 5:28  
136.0 C 229kPa 5:58  
136.2 C 232kPa 6:28  
136.4 C 233kPa 6:58  
136.2 C 229kPa 7:28  
136.3 C 232kPa 7:58  
136.3 C 232kPa 7:58  
VENTING 7:58  
AIR DRYING 8:52  
CYCLE COMPLETE 8:59  
-----

Digital Signature #  
CB78FEAD1E027923  
-----