
Midmark M11 - v1.0.5
Total Cycles: 1277

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.7 F 0.0
2:00 136.2 F 1.4
4:00 161.1 F 4.5
6:00 184.5 F 8.6
8:00 202.1 F 13.0
10:00 217.7 F 5.9
12:00 229.1 F 10.4
14:00 232.4 F 8.0
16:00 241.0 F 12.2
18:00 239.9 F 10.7
20:00 252.4 F 17.0
22:00 263.2 F 23.5
24:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.2 F 29.8
2:00 272.5 F 29.5
3:00 271.9 F 29.1
4:00 271.1 F 28.8

Min 271.1 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 238.0 F 0.0
10:00 201.0 F 0.0
15:00 188.1 F 0.0
20:00 186.4 F 0.0
25:00 183.2 F 0.0
30:00 181.9 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

07/08/2022 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____