
Midmark M11 - v1.0.0
Total Cycles: 3672

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 178.1 F 3.8
4:00 187.8 F 8.3
6:00 209.2 F 12.3
8:00 223.2 F 7.8
10:00 237.1 F 13.0
12:00 237.2 F 10.2
14:00 235.1 F 8.0
16:00 249.5 F 15.2
18:00 261.7 F 22.1
20:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.1 F 28.7
2:00 272.5 F 29.2
3:00 271.7 F 28.7
4:00 272.2 F 28.9

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 224.9 F 0.0
10:00 232.0 F 0.0
15:00 229.0 F 0.0
20:00 230.7 F 0.0
25:00 227.4 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:00
CYCLE COMPLETE

08/17/2023 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____