

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4360

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2024 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.3 F 0.0  
2:00 129.9 F 0.7  
4:00 150.3 F 2.4  
6:00 176.3 F 5.1  
8:00 175.5 F 8.5  
10:00 191.3 F 12.3  
12:00 210.6 F 4.1  
14:00 224.9 F 7.9  
16:00 236.2 F 12.0  
18:00 232.7 F 8.1  
20:00 243.5 F 13.1  
22:00 241.4 F 11.1  
24:00 252.9 F 16.8  
26:00 263.2 F 22.8  
28:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.8  
1:00 272.6 F 29.3  
2:00 271.8 F 29.1  
3:00 272.3 F 28.9  
4:00 272.5 F 29.4

Min 271.5 F 28.7  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.6 F 0.0  
5:00 227.0 F 0.0  
10:00 226.6 F 0.0  
15:00 229.7 F 0.0  
20:00 228.0 F 0.0  
25:00 229.0 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:57  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:07:35  
CYCLE COMPLETE

03/06/2024 01:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_