
Midmark M11 - v1.0.0
Total Cycles: 4039

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2023 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 183.6 F 1.9
4:00 184.0 F 5.3
6:00 208.6 F 9.6
8:00 222.6 F 7.2
10:00 225.7 F 8.0
12:00 236.1 F 12.7
14:00 236.2 F 9.5
16:00 239.7 F 9.6
18:00 247.0 F 13.5
20:00 258.2 F 19.5
22:00 268.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.6 F 29.0
2:00 272.4 F 28.7
3:00 272.5 F 29.1
4:00 271.8 F 28.5

Min 271.0 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 209.6 F 0.0
10:00 212.8 F 0.0
15:00 213.4 F 0.0
20:00 214.9 F 0.0
25:00 220.3 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

11/20/2023 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____