

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1120

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2022 08:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.4 F 0.0  
2:00 172.5 F 3.0  
4:00 190.5 F 7.7  
6:00 213.5 F 13.0  
8:00 224.9 F 8.4  
10:00 236.9 F 11.7  
12:00 240.0 F 11.8  
14:00 239.7 F 10.7  
16:00 253.1 F 17.4  
18:00 264.6 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.5  
2:00 271.4 F 29.1  
3:00 272.5 F 29.6  
4:00 271.5 F 28.9  
  
Min 271.4 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
5:00 182.1 F 0.0  
10:00 179.2 F 0.0  
15:00 180.2 F 0.0  
20:00 162.5 F 0.0  
25:00 157.6 F 0.0  
30:00 152.5 F 0.0  
Drying Complete

FILLING: 2:11  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

05/26/2022 09:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02