
Midmark M11 - v1.0.0
Total Cycles: 3408

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.6 F 0.0
2:00 187.4 F 3.5
4:00 195.6 F 7.7
6:00 218.7 F 12.8
8:00 227.7 F 7.8
10:00 239.7 F 13.1
12:00 238.7 F 10.5
14:00 237.4 F 8.6
16:00 250.6 F 15.5
18:00 262.1 F 22.1
20:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.5 F 29.1
2:00 271.6 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.3

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 209.6 F 0.0
10:00 218.6 F 0.0
15:00 229.1 F 0.0
20:00 232.4 F 0.0
25:00 226.9 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

06/19/2023 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____