
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003805
9:42 09/06/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:24
STERILIZING 3:39
136.5 C 222kPa 3:39
136.4 C 225kPa 4:09
136.3 C 224kPa 4:39
136.5 C 224kPa 5:09
136.4 C 220kPa 5:39
136.4 C 224kPa 6:09
136.1 C 218kPa 6:39
136.3 C 224kPa 7:09
136.3 C 224kPa 7:09
VENTING 7:09
AIR DRYING 8:22
CYCLE COMPLETE 8:32

Digital Signature #
730B5E04BF22DF3E
