

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000258  
12:31 30/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:37  
PRESSURIZING 3:46  
STERILIZING 5:26  
136.4 C 225kPa 5:26  
136.5 C 233kPa 5:56  
136.3 C 232kPa 6:26  
136.3 C 233kPa 6:56  
136.3 C 232kPa 7:26  
136.4 C 233kPa 7:56  
136.3 C 233kPa 8:26  
136.2 C 231kPa 8:56  
136.2 C 231kPa 8:56  
VENTING 8:56  
AIR DRYING 9:57  
CYCLE COMPLETE 24:57  
-----

Digital Signature #  
FA91F76D029EF083  
-----