
Midmark M11 - v1.0.0
Total Cycles: 1693

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2022 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 161.7 F 1.5
4:00 180.0 F 4.9
6:00 199.6 F 9.4
8:00 219.4 F 6.6
10:00 227.5 F 8.6
12:00 237.9 F 12.0
14:00 240.2 F 11.2
16:00 238.0 F 9.5
18:00 252.2 F 16.4
20:00 263.9 F 23.3
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.3
3:00 272.5 F 29.0
4:00 271.8 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 208.1 F 0.0
10:00 221.3 F 0.0
15:00 228.8 F 0.0
20:00 227.6 F 0.0
25:00 229.7 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

03/02/2022 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____