

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4335

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2024 06:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 105.0 F 0.0  
2:00 105.5 F 1.0  
4:00 121.8 F 3.3  
6:00 152.9 F 6.4  
8:00 188.9 F 10.1  
10:00 206.3 F 3.9  
12:00 214.2 F 6.5  
14:00 225.9 F 10.5  
16:00 230.5 F 7.4  
18:00 239.3 F 11.2  
20:00 235.8 F 8.6  
22:00 248.9 F 14.7  
24:00 259.5 F 20.6  
26:00 269.9 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.1 F 28.6  
2:00 271.9 F 28.8  
3:00 271.8 F 28.4  
4:00 272.3 F 29.0

Min 271.3 F 28.2  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 215.0 F 0.0  
10:00 220.9 F 0.0  
15:00 228.4 F 0.0  
20:00 230.7 F 0.0  
25:00 229.5 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 26:53  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:05:25  
CYCLE COMPLETE

02/26/2024 07:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_