
Midmark M11 - v1.0.5
Total Cycles: 4337

Name: RITTER2
Sterilizer ID: V2313228
02/06/2025 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.4 F 0.0
4:00 74.0 F 0.5
6:00 98.8 F 2.3
8:00 127.1 F 5.0
10:00 155.8 F 8.1
12:00 173.9 F 11.4
14:00 212.7 F 3.7
16:00 221.5 F 6.3
18:00 232.1 F 9.8
20:00 233.8 F 12.0
22:00 235.9 F 9.9
24:00 238.6 F 10.0
26:00 244.5 F 13.0
28:00 255.0 F 18.7
30:00 265.1 F 24.6
32:00 270.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.8
1:00 272.1 F 30.2
2:00 272.8 F 30.1
3:00 273.0 F 29.9
4:00 272.5 F 29.8

Min 271.5 F 29.7
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 224.7 F 0.0
10:00 212.6 F 0.0
15:00 202.8 F 0.0
20:00 198.0 F 0.0
25:00 193.1 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 32:32
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:11:08
CYCLE COMPLETE

02/06/2025 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____