
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001843
8:41 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:19
STERILIZING 3:25
136.5 C 226kPa 3:25
136.3 C 224kPa 3:55
136.4 C 225kPa 4:25
136.4 C 225kPa 4:55
136.2 C 223kPa 5:25
136.4 C 225kPa 5:55
136.4 C 226kPa 6:25
136.4 C 225kPa 6:55
136.4 C 225kPa 6:55
VENTING 6:55
AIR DRYING 7:48
CYCLE COMPLETE 8:02

Digital Signature #
B3AD5FD1C549929F
