

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001529  
13:10 23/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:38  
STERILIZING 3:40  
136.0 C 223kPa 3:40  
136.2 C 224kPa 4:10  
136.4 C 226kPa 4:40  
136.4 C 226kPa 5:10  
136.4 C 224kPa 5:40  
136.4 C 224kPa 6:10  
136.4 C 224kPa 6:40  
136.4 C 226kPa 7:10  
136.4 C 226kPa 7:10  
VENTING 7:10  
AIR DRYING 8:04  
CYCLE COMPLETE 8:33  
-----

Digital Signature #  
084E7180D2CEE4F1  
-----