
Midmark M11 - v1.0.0
Total Cycles: 4128

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 75.8 F 0.3
4:00 88.8 F 1.2
6:00 123.3 F 3.6
8:00 154.4 F 6.4
10:00 172.8 F 9.9
12:00 193.7 F 11.0
14:00 209.4 F 4.0
16:00 218.0 F 7.0
18:00 228.0 F 10.3
20:00 231.9 F 9.2
22:00 236.6 F 9.9
24:00 237.8 F 9.6
26:00 244.5 F 12.7
28:00 255.7 F 18.6
30:00 265.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.2 F 29.0
2:00 271.9 F 29.0
3:00 271.9 F 28.7
4:00 272.2 F 29.3

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 196.6 F 0.0
10:00 191.5 F 0.0
15:00 185.8 F 0.0
20:00 181.7 F 0.0
25:00 179.0 F 0.0
30:00 177.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:48
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

12/18/2023 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____