

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001211  
13:19 11/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:31  
PRESSURIZING 4:32  
STERILIZING 5:44  
136.4 C 223kPa 5:44  
136.4 C 225kPa 6:14  
136.3 C 225kPa 6:44  
136.4 C 225kPa 7:14  
136.1 C 222kPa 7:44  
136.3 C 223kPa 8:14  
136.3 C 225kPa 8:44  
136.1 C 223kPa 9:14  
136.1 C 223kPa 9:14  
VENTING 9:14  
AIR DRYING 10:11  
CYCLE COMPLETE 25:11  
-----

Digital Signature #  
CF217F6576DB7A7F  
-----