

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000055  
8:35 29/01/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:10  
STERILIZING 4:20  
136.4 C 223kPa 4:20  
136.2 C 223kPa 4:50  
136.2 C 222kPa 5:20  
136.2 C 221kPa 5:50  
136.1 C 222kPa 6:20  
136.4 C 223kPa 6:50  
136.2 C 222kPa 7:20  
136.4 C 224kPa 7:50  
136.4 C 224kPa 7:50  
VENTING 7:50  
AIR DRYING 8:41  
CYCLE COMPLETE 12:21

-----  
Digital Signature #  
A21B68C1D43AFE8C  
-----