
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000594
12:12 20/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:11
STERILIZING 3:21
136.0 C 227kPa 3:21
136.3 C 231kPa 3:51
136.3 C 232kPa 4:21
136.4 C 233kPa 4:51
136.4 C 231kPa 5:21
136.3 C 232kPa 5:51
136.0 C 230kPa 6:21
136.3 C 232kPa 6:51
136.3 C 232kPa 6:51
VENTING 6:51
AIR DRYING 7:38
CYCLE COMPLETE 22:38

Digital Signature #
E8806F22FE5388B6
