
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001674
12:51 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:33
136.0 C 220kPa 4:33
136.2 C 224kPa 5:03
136.2 C 224kPa 5:33
136.4 C 223kPa 6:03
136.4 C 223kPa 6:33
136.3 C 223kPa 7:03
136.1 C 223kPa 7:33
136.4 C 224kPa 8:03
136.4 C 224kPa 8:03
VENTING 8:03
AIR DRYING 8:58
CYCLE COMPLETE 9:07

Digital Signature #
A53EE1C618D5CB02
