

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4101

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2023 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 69.8 F 0.2  
4:00 77.4 F 1.1  
6:00 109.4 F 3.4  
8:00 146.3 F 6.3  
10:00 172.6 F 9.7  
12:00 191.2 F 13.3  
14:00 206.6 F 3.8  
16:00 218.9 F 6.9  
18:00 228.2 F 10.4  
20:00 233.4 F 8.5  
22:00 238.3 F 10.2  
24:00 236.5 F 8.7  
26:00 246.1 F 13.2  
28:00 257.0 F 19.1  
30:00 267.0 F 25.2  
32:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.6  
1:00 272.6 F 28.8  
2:00 272.7 F 29.0  
3:00 272.5 F 29.1  
4:00 272.0 F 29.0  
  
Min 271.5 F 28.5  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.9 F 0.0  
5:00 245.0 F 0.0  
10:00 226.0 F 0.0  
15:00 227.6 F 0.0  
20:00 227.6 F 0.0  
25:00 227.2 F 0.0  
30:00 227.6 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 32:05  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:10:26  
CYCLE COMPLETE

12/11/2023 10:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_