
Midmark M11 - v1.0.5
Total Cycles: 1905

Name: RITTER2
Sterilizer ID: V2313228
01/09/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 178.1 F 2.6
4:00 175.0 F 6.2
6:00 197.8 F 11.0
8:00 214.9 F 5.4
10:00 228.0 F 10.8
12:00 230.5 F 7.7
14:00 242.1 F 12.1
16:00 243.9 F 12.9
18:00 256.2 F 19.5
20:00 267.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.6 F 30.2
2:00 272.1 F 29.5
3:00 271.5 F 29.5
4:00 272.6 F 30.2

Min 271.5 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 217.7 F 0.0
10:00 237.3 F 0.0
15:00 233.0 F 0.0
20:00 226.3 F 0.0
25:00 228.2 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

01/09/2023 12:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____