
Midmark M11 - v1.0.5
Total Cycles: 1

Name: RITTER2
Sterilizer ID: V2313228
04/21/2021 07:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.3 F 0.0
2:00 195.2 F 2.0
4:00 221.4 F 9.5
6:00 232.2 F 8.8
8:00 237.7 F 9.6
10:00 240.9 F 11.0
12:00 257.7 F 19.8
14:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 271.9 F 28.8
2:00 271.9 F 28.9
3:00 271.9 F 28.8
4:00 271.9 F 28.6

Min 271.8 F 28.5
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
Dry Cycle Aborted

FILLING: 4:03
HEATING: 14:30
STERILIZING: 4:00
VENTING: 2:50
DRYING: 0:26
TOTAL CYCLE: 00:25:49

04/21/2021 07:28 AM
ACCEPTED: X
REJECTED:
Operator: U01