
Midmark M11 - v1.0.0
Total Cycles: 808

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2021 11:42 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 189.8 F 3.8
4:00 208.2 F 2.7
6:00 221.9 F 5.7
8:00 223.6 F 4.1
10:00 228.3 F 5.2
12:00 238.8 F 9.7
14:00 251.3 F 15.8

STERILIZING
mm:ss Degrees PSI
0:00 251.7 F 15.9
1:00 252.3 F 16.0
2:00 252.9 F 16.4
3:00 251.9 F 15.8
4:00 252.9 F 16.4
5:00 252.0 F 15.8
6:00 252.9 F 16.5
7:00 252.4 F 16.1
8:00 251.9 F 15.9
9:00 252.9 F 16.4
10:00 252.5 F 16.1
11:00 252.0 F 15.8
12:00 252.7 F 16.4
13:00 252.9 F 16.4
14:00 252.5 F 16.1
15:00 252.0 F 15.9
16:00 252.1 F 16.1
17:00 253.1 F 16.5
18:00 252.8 F 16.3
19:00 252.4 F 16.1
20:00 252.1 F 15.9
21:00 251.8 F 15.9
22:00 252.8 F 16.5
23:00 253.0 F 16.5
24:00 252.7 F 16.3
25:00 252.4 F 16.1
26:00 252.0 F 15.9
27:00 251.8 F 15.9
28:00 252.8 F 16.5
29:00 253.0 F 16.5
30:00 252.8 F 16.3

Min 251.7 F 15.8
Max 253.2 F 16.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 186.3 F 0.0
10:00 192.5 F 0.0
15:00 208.6 F 0.0
20:00 208.1 F 0.0
25:00 206.9 F 0.0
30:00 169.0 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 14:35
STERILIZING: 30:00

VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:19:38
CYCLE COMPLETE

04/26/2021 01:02 PM
ACCEPTED: X
REJECTED:
Operator: U01