

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2945

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2023 10:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
2:00 192.7 F 3.1  
4:00 217.1 F 8.3  
6:00 233.7 F 9.6  
8:00 234.6 F 10.0  
10:00 237.6 F 8.1  
12:00 245.6 F 12.8  
14:00 244.8 F 12.6  
16:00 257.8 F 19.5  
18:00 269.2 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.6 F 29.1  
2:00 271.6 F 28.2  
3:00 272.6 F 29.1  
4:00 271.7 F 28.3  
  
Min 271.3 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 236.4 F 0.0  
10:00 226.7 F 0.0  
15:00 227.6 F 0.0  
20:00 225.0 F 0.0  
25:00 224.0 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:01  
CYCLE COMPLETE

02/27/2023 11:43 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_