
Midmark M11 - v1.0.0
Total Cycles: 1172

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2021 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 198.7 F 2.5
4:00 199.1 F 7.2
6:00 230.1 F 13.5
8:00 223.1 F 9.5
10:00 232.8 F 8.3
12:00 240.9 F 13.0
14:00 241.7 F 11.6
16:00 255.1 F 18.5
18:00 267.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 273.0 F 29.6
2:00 271.7 F 28.9
3:00 272.3 F 29.5
4:00 272.8 F 29.5

Min 271.2 F 28.8
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 237.0 F 0.0
10:00 228.4 F 0.0
15:00 228.1 F 0.0
20:00 230.0 F 0.0
25:00 233.0 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

09/14/2021 04:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____