
Midmark M11 - v1.0.0
Total Cycles: 4691

Name: RITTER 1
Sterilizer ID: V2178198
07/02/2024 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 74.7 F 0.2
4:00 87.5 F 1.1
6:00 121.3 F 3.4
8:00 149.5 F 6.3
10:00 172.3 F 9.4
12:00 190.8 F 12.9
14:00 205.7 F 3.4
16:00 215.9 F 6.6
18:00 226.1 F 10.0
20:00 232.1 F 10.0
22:00 235.7 F 9.8
24:00 238.3 F 9.4
26:00 245.1 F 12.7
28:00 256.5 F 18.7
30:00 266.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 28.8
2:00 271.1 F 28.1
3:00 272.4 F 28.8
4:00 271.3 F 28.1

Min 271.1 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 206.1 F 0.0
10:00 196.8 F 0.0
15:00 195.0 F 0.0
20:00 198.3 F 0.0
25:00 202.2 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:48
CYCLE COMPLETE

07/02/2024 10:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____