

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2588

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2022 03:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
2:00 206.3 F 5.2  
4:00 212.5 F 10.9  
6:00 223.9 F 6.6  
8:00 237.6 F 12.5  
10:00 238.8 F 10.5  
12:00 237.0 F 9.2  
14:00 252.1 F 16.4  
16:00 264.3 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.2  
3:00 272.3 F 29.1  
4:00 272.1 F 28.7  
  
Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 233.5 F 0.0  
10:00 229.5 F 0.0  
15:00 229.1 F 0.0  
Dry Cycle Aborted

FILLING: 2:22  
HEATING: 17:55  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 18:50  
TOTAL CYCLE: 00:46:21

11/21/2022 04:13 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_