
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005008
10:22 30/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:54
STERILIZING 3:54
136.5 C 223kPa 3:54
136.1 C 221kPa 4:24
136.4 C 224kPa 4:54
136.4 C 223kPa 5:24
136.4 C 223kPa 5:54
136.0 C 217kPa 6:24
136.4 C 223kPa 6:54
136.4 C 223kPa 7:24
136.4 C 223kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 11:36

Digital Signature #
2FDE62A2232E0D8A
