
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000752
10:06 02/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:55
136.2 C 220kPa 4:55
136.4 C 225kPa 5:25
136.4 C 224kPa 5:55
136.4 C 225kPa 6:25
136.3 C 223kPa 6:55
136.4 C 225kPa 7:25
136.4 C 224kPa 7:55
136.4 C 225kPa 8:25
136.4 C 225kPa 8:25
VENTING 8:25
AIR DRYING 9:21
CYCLE COMPLETE 9:24

Digital Signature #
63081D1C2D4A5C7B
