
Midmark M11 - v1.0.5
Total Cycles: 3182

Name: RITTER2
Sterilizer ID: V2313228
12/04/2023 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.0 F 0.0
2:00 142.8 F 1.0
4:00 155.5 F 3.9
6:00 182.0 F 7.8
8:00 202.2 F 11.9
10:00 222.2 F 6.3
12:00 226.6 F 10.9
14:00 230.7 F 7.3
16:00 242.9 F 12.8
18:00 241.6 F 11.6
20:00 254.0 F 17.9
22:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 30.0
2:00 272.4 F 30.1
3:00 271.6 F 29.5
4:00 272.4 F 29.6

Min 271.0 F 28.9
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 219.3 F 0.0
10:00 209.3 F 0.0
15:00 203.6 F 0.0
20:00 208.1 F 0.0
25:00 221.2 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

12/04/2023 05:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____