
Midmark M11 - v1.0.0
Total Cycles: 1022

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 172.3 F 3.1
4:00 188.9 F 7.2
6:00 211.9 F 12.3
8:00 224.6 F 7.0
10:00 238.8 F 12.1
12:00 236.5 F 9.4
14:00 238.8 F 9.4
16:00 248.4 F 14.4
18:00 260.3 F 21.0
20:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.1
4:00 272.0 F 28.5

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 226.9 F 0.0
Dry Cycle Aborted

FILLING: 1:34
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 9:57
TOTAL CYCLE: 00:39:25

07/15/2021 03:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____