
Midmark M11 - v1.0.5
Total Cycles: 4239

Name: RITTER2
Sterilizer ID: V2313228
01/02/2025 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.5 F 0.0
2:00 127.5 F 0.4
4:00 134.8 F 2.4
6:00 171.0 F 6.0
8:00 188.9 F 10.1
10:00 217.1 F 7.6
12:00 221.5 F 8.9
14:00 236.4 F 12.3
16:00 236.4 F 10.4
18:00 235.3 F 8.6
20:00 248.0 F 14.8
22:00 259.2 F 21.2
24:00 269.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 271.7 F 29.8
2:00 272.8 F 30.4
3:00 272.9 F 30.2
4:00 272.4 F 30.0

Min 271.1 F 29.1
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 231.6 F 0.0
10:00 223.8 F 0.0
15:00 217.6 F 0.0
20:00 214.0 F 0.0
25:00 211.2 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:51
CYCLE COMPLETE

01/02/2025 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____