

-----  
Midmark M11 - v1.0.0  
Total Cycles: 933

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2021 03:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
2:00 187.7 F 2.6  
4:00 198.4 F 6.2  
6:00 211.9 F 10.4  
8:00 221.6 F 7.2  
10:00 234.7 F 9.7  
12:00 235.7 F 9.8  
14:00 242.4 F 11.9  
16:00 240.1 F 10.3  
18:00 252.3 F 16.2  
20:00 262.6 F 22.2  
22:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.4 F 28.7  
2:00 271.8 F 28.4  
3:00 272.6 F 28.9  
4:00 271.8 F 28.7

Min 271.0 F 28.0  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 228.4 F 0.0  
10:00 200.6 F 0.0  
15:00 195.1 F 0.0  
20:00 187.2 F 0.0  
25:00 194.3 F 0.0  
30:00 207.0 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:37  
CYCLE COMPLETE

06/23/2021 04:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_