
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008316
10:01 13/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:13
STERILIZING 4:04
136.0 C 220kPa 4:04
136.2 C 223kPa 4:34
136.3 C 223kPa 5:04
136.4 C 224kPa 5:34
136.4 C 224kPa 6:04
136.4 C 225kPa 6:34
136.4 C 224kPa 7:04
136.4 C 223kPa 7:34
136.4 C 223kPa 7:34
VENTING 7:34
AIR DRYING 8:27
CYCLE COMPLETE 8:58

Digital Signature #
3A4D40439E9EF788
