
Midmark M11 - v1.0.0
Total Cycles: 1410

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 200.2 F 3.3
4:00 214.2 F 8.3
6:00 220.2 F 9.8
8:00 234.3 F 9.7
10:00 231.9 F 8.4
12:00 244.9 F 13.0
14:00 242.8 F 11.7
16:00 255.8 F 18.4
18:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.7
2:00 271.7 F 28.6
3:00 272.4 F 28.7
4:00 271.5 F 28.5

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 205.4 F 0.0
10:00 196.6 F 0.0
15:00 196.6 F 0.0
20:00 195.0 F 0.0
25:00 193.7 F 0.0
30:00 194.0 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

11/30/2021 04:17 PM
ACCEPTED: X
REJECTED:
Operator: U02