

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5188

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2024 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.3 F 0.0  
2:00 109.9 F 0.0  
4:00 137.1 F 0.0  
6:00 164.9 F 0.1  
8:00 184.5 F 0.3  
10:00 206.1 F 2.6  
12:00 221.6 F 6.3  
14:00 230.0 F 10.1  
16:00 236.1 F 9.1  
18:00 240.1 F 10.6  
20:00 236.3 F 8.0  
22:00 245.3 F 12.0  
24:00 246.4 F 12.8  
26:00 258.0 F 19.1  
28:00 267.9 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.2 F 28.3  
2:00 272.8 F 28.5  
3:00 272.4 F 28.2  
4:00 271.9 F 27.8

Min 271.6 F 27.7  
Max 272.8 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
5:00 228.4 F 0.0  
10:00 221.9 F 0.0  
15:00 209.7 F 0.0  
20:00 195.8 F 0.0  
25:00 202.6 F 0.0  
30:00 198.6 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 29:32  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:07:56  
CYCLE COMPLETE

12/13/2024 04:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_