

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1902

Name: RITTER2  
Sterilizer ID: V2313228  
01/05/2023 03:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.1 F 0.0  
2:00 160.7 F 3.2  
4:00 177.6 F 7.1  
6:00 199.8 F 12.0  
8:00 216.9 F 6.2  
10:00 232.5 F 11.3  
12:00 232.5 F 8.4  
14:00 241.3 F 11.1  
16:00 245.0 F 13.3  
18:00 257.4 F 20.0  
20:00 268.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.3  
1:00 272.0 F 29.9  
2:00 272.3 F 29.6  
3:00 271.5 F 29.4  
4:00 272.4 F 30.0

Min 271.4 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 231.7 F 0.0  
10:00 228.6 F 0.0  
15:00 229.2 F 0.0  
20:00 227.8 F 0.0  
25:00 228.3 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:09  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:34  
CYCLE COMPLETE

01/05/2023 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_