
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001851
13:26 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:01
STERILIZING 4:09
136.0 C 222kPa 4:09
136.3 C 225kPa 4:39
136.4 C 226kPa 5:09
136.4 C 225kPa 5:39
136.2 C 225kPa 6:09
136.4 C 225kPa 6:39
136.3 C 225kPa 7:09
136.4 C 224kPa 7:39
136.4 C 224kPa 7:39
VENTING 7:39
AIR DRYING 8:33
CYCLE COMPLETE 11:06

Digital Signature #
BA17B65452587FE8
