

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2688

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2022 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.9 F 0.0  
2:00 76.0 F 0.0  
4:00 82.0 F 0.0  
6:00 106.0 F 0.5  
8:00 137.9 F 2.3  
10:00 167.4 F 4.9  
12:00 188.0 F 8.3  
14:00 203.1 F 12.1  
16:00 217.9 F 4.9  
18:00 228.4 F 9.0  
20:00 237.6 F 13.4  
22:00 237.7 F 10.2  
24:00 236.1 F 8.3  
26:00 248.0 F 14.2  
28:00 259.3 F 20.4  
30:00 269.6 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.3 F 28.8  
2:00 271.7 F 28.7  
3:00 272.1 F 28.7  
4:00 271.8 F 28.8

Min 271.4 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 218.1 F 0.0  
10:00 200.7 F 0.0  
15:00 218.4 F 0.0  
20:00 216.3 F 0.0  
25:00 221.2 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:58  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:60  
CYCLE COMPLETE

12/19/2022 10:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_