

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3537

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 03:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.8 F 0.0  
2:00 147.4 F 0.6  
4:00 170.4 F 2.6  
6:00 192.8 F 5.6  
8:00 206.1 F 9.8  
10:00 222.2 F 9.4  
12:00 226.9 F 7.8  
14:00 237.6 F 12.3  
16:00 236.5 F 9.6  
18:00 238.5 F 9.6  
20:00 247.5 F 14.0  
22:00 259.1 F 20.3  
24:00 269.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.5 F 29.0  
2:00 271.3 F 28.2  
3:00 272.2 F 29.1  
4:00 272.0 F 28.7

Min 271.0 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 216.2 F 0.0  
10:00 220.3 F 0.0  
15:00 228.7 F 0.0  
20:00 217.3 F 0.0  
25:00 211.6 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 25:05  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:03:41  
CYCLE COMPLETE

07/17/2023 04:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_