
Midmark M11 - v1.0.5
Total Cycles: 2743

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 186.9 F 3.0
4:00 193.6 F 7.5
6:00 211.5 F 12.7
8:00 223.8 F 7.6
10:00 236.3 F 12.6
12:00 236.6 F 10.1
14:00 236.3 F 8.9
16:00 248.1 F 14.8
18:00 259.2 F 21.0
20:00 269.4 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.8 F 28.9
3:00 272.1 F 29.6
4:00 272.5 F 29.5

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 211.6 F 0.0
10:00 223.9 F 0.0
15:00 222.9 F 0.0
20:00 221.0 F 0.0
25:00 222.9 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

08/09/2023 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____