
Midmark M11 - v1.0.5
Total Cycles: 2667

Name: RITTER2
Sterilizer ID: V2313228
07/19/2023 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 175.9 F 2.7
4:00 193.4 F 7.4
6:00 211.3 F 12.9
8:00 225.4 F 8.3
10:00 237.5 F 12.1
12:00 240.0 F 11.6
14:00 239.1 F 10.4
16:00 252.4 F 17.1
18:00 264.0 F 23.9
20:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.4
4:00 272.6 F 29.7

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 228.8 F 0.0
10:00 228.0 F 0.0
15:00 227.5 F 0.0
20:00 225.0 F 0.0
25:00 223.3 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

07/19/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____