
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001932
11:16 08/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:03
136.4 C 224kPa 3:03
136.3 C 225kPa 3:33
136.3 C 226kPa 4:03
136.0 C 222kPa 4:33
136.4 C 225kPa 5:03
136.4 C 226kPa 5:33
136.4 C 226kPa 6:03
136.2 C 223kPa 6:33
136.2 C 223kPa 6:33
VENTING 6:33
AIR DRYING 7:24
CYCLE COMPLETE 37:24

Digital Signature #
21522D424137E335
